

#4

Fig. 1





Fig. 2A

Fig. 2B




Fig.3A

<i>Fig.3A-1</i>	<i>Fig.3A-2</i>
<i>Fig.3A-3</i>	<i>Fig.3A-4</i>

Fig.3B

<i>Fig.3B-1</i>	<i>Fig.3B-2</i>
<i>Fig.3B-3</i>	<i>Fig.3B-4</i>

Fig.3B-1





Fig.3B-2



3B-3

WASTE BAG (134b)

TRANSFER BAG (197)

HEMO INLET PUMP (37)

HEMO-CONCENTRATOR

VENOUS ENTRY MODULE (108)

PRE-DONATION PORT

PRE-BYPASS FILTER (168)

ARTERIAL VENOUS SHUNT (V30, 93)

CARDIOPLEGIA BLOOD ACCESS (119)

ARTERIAL PATIENT LINE (122)

BUBBLE SENSOR (126)

ARTERIAL FILTER (118a)

Pressure points: P, P1, P6 (Aux. Pressure), P9

Flow indicators: 134, 134a, 130, 14, 104, 166, 119, 94, 118a, 116b

Valves: V?, V31 (45), V30 (93), V27 (92)

Sensors: T2 (88)

Other components: 86, 197, 104, 175, 124, 116b

Fig.3B-4





Fig. 3A-1

 = STOPCOCK

 = AUTOMATIC VALVE





Fig.3A-2

Fig.3A-3





Fig.3A-4


Fig. 4

Fig. 5A




Fig. 5B




Fig. 5D




Fig. 5C




Fig. 5E




Fig. 5F




Fig. 6A





Fig. 6B

Fig. 6C




Fig. 6D




Fig. 6E






Fig. 7A*Fig. 7B*


Fig. 8A

Fig. 8B




Fig. 8C




Fig. 8D




Fig. 8E










Fig. 9A*Fig. 9B**Fig. 9C**Fig. 9D**Fig. 9E**Fig. 9F*

Fig. 10B




Fig. 10A






Fig. 10D**Fig. 10C**

Fig. 11




Fig. 12




Fig. 13




Fig. 14




Fig. 15




Fig. 16




Fig. 17




Fig. 18





Fig. 19A

Fig. 19B





Fig. 20A

Fig. 20B





Fig. 21

Fig. 22




Fig.23A




Fig.23B




Fig. 24A





Fig. 24B



The diagram illustrates a pump control system 10, which is a closed-loop system for managing fluid flow. The system is composed of several interconnected components:

- External Systems:** These provide input to the **Embedded Processor 300**.
- Pump User Interface:** This interface is connected to the **Embedded Processor 300** and the **Pump Control/Monitor Processor 312**.
- Temperature Control Systems 330:** These systems are connected to the **Embedded Processor 300** and the **Pump Control/Monitor Processor 312**.
- Gas Sources:** These provide input to the **Gas Circuit 340**.
- Control Unit 10:** This unit contains the **Embedded Processor 300**, the **Pump Control/Monitor Processor 312**, and the **Control/Monitor Processors 306**.
- Disposable Assembly 100:** This assembly is the central component, containing **Pumps 31-38**, **VLC 46**, and **Valve Assemblies**. It is connected to the **Gas Circuit 340** and the **Control/Monitor Processors 306**.
- Pressure, Temperature, Bubble, Sensors 310:** These sensors are connected to the **Control/Monitor Processors 306** and provide feedback to the **Embedded Processor 300** and the **Pump Control/Monitor Processor 312**.
- User Interface 50:** This interface includes a **GUI Display 54** and a **Control Knob 52**, connected to the **Embedded Processor 300** and the **Pump Control/Monitor Processor 312**.
- Flow Lines:** The system includes several flow lines: **Cardiology Mix 156**, **Suction Line 1 170**, **Suction Line 2 172**, **Vent Line 186**, **Venous Blood 104**, and **Return Blood 122**.

The system is designed to provide precise control over fluid flow, ensuring accurate dosing and monitoring of the patient's condition.

Fig. 26




Fig. 27









VENOUS 		ARTERIAL LINE PRESSURE 0mmHg FLOW 0.00 lpm TEMP. 0.0°C		CARDIOPLEGIA LINE PRESSURE 0mmHg FLOW BOLUS 0.000 lpm omi TEMP. 0.0°C ISCHEMIC TIME 00:00		BLENDER FLOW 3.2 lpm FIO2 50% FICO2 0%		OTHER PATIENT ARTERIAL 110 mmHg PATIENT TEMP. 32.0°C FECO2 1% 8:47	
FOLLOW INSTRUCTIONS AND THEN PRESS 'LOAD' TO GO TO LOAD SCREEN.									
USER SET UP		1.  INSERT OXYGENATOR AND VENOUS RESERVOIR IN HOLDERS.							
A-V		2.  SNAP IN PRE-BYPASS FILTER AND VENOUS ENTRY MODULE IN HOLDERS AND PLACE VENOUS LINE IN CLAMP AND CLOSE COVER.							
CPG		3.  INSERT CARTRIDGE AND ARTERIAL FILTER IN HOLDERS.							
SUCTION/FLUIDS		4.  CONNECT LINES TO ARTERIAL FILTER, VENOUS ENTRY MODULE, AND VENOUS RESERVOIR(2).							
GASES		5.  PLACE LINE IN BUBBLE SENSOR AND CLOSE COVER.							
WAVEFORMS		6.  PLACE ARTERIAL AND CARDIOPLEGIA TABLE LINES IN CLAMPS AND CLOSE COVERS.							
SETTINGS		7.  INSTALL PUMP LOOPS AND CLOSE ALL LIDS. HANG TABLE PACK ON CONSOLE.							
UNLOAD		LOAD							

Fig.28A




Fig.28B




Fig.28C




Fig. 28D




Fig. 28E




Fig. 28F

208b


LOW LEVEL-PUMP STOPPED		RESTART PUMP	ARTERIAL PRESSURE LIMIT	RESTART PUMP	MUTE
ARTERIAL BUBBLE		RESTART PUMP	ARTERIAL HIGH TEMPERATURE	LESS	
VENOUS		ARTERIAL PRESSURE LOW			
SAT 75%		CRYSTAL POT MUST BE OFF			
HCT 25%		PLEAGIA K + > 40 mEq/L			
TEMP 32°C		ON BATTERY			
		ARTERIAL PRESSURE 200mmHg FLOW 0.00 lpm TEMP 32°C			
<input type="button" value="FULL OPEN"/> <input type="button" value="0%"/> <input type="button" value="FULL CLOSE"/>					

Fig. 29




Fig. 30A




Fig. 30B




Fig. 30C




Fig. 30D




Fig. 30E




Fig. 30F




Fig. 30G




Fig. 30H




Fig. 30I




Fig. 30J

MAIN	FILL PATIENT: START ARTERIAL TO FLOW DOWN PATIENT LINE.									
A-V	SYSTEM									
CPG	PRE-BYPASS FILTER		SYSTEM RECIRC.		TEST ARTERIAL CONNECTION		PATIENT INFO.		LOG EVENT	
SUCTION/FLUIDS	USER DEFINED: 264d' 264d"									
GASES	FILL PATIENT		CHASE		TO BAGS					
WAVEFORMS	BOLUS ml	DELIVER 0 ml	BOLUS ml	DELIVER ml						
SETTINGS	TIMERS:									
	ON-BYPASS 0:05:57	RESET	OFF-BYPASS 00:45	RESET	X-CLAMP 0:06:34	RESET	AUXILIARY 0:03:47	RESET		
									RETURN TO < BYPASS	MOVE TO UNLOADING

Fig.30K




Fig.30L




Fig. 31A






Fig. 31B*Fig. 31C*




Fig.31D*Fig.31E**Fig.31F*

Fig. 32A




Fig.32B




Fig.32C




Fig. 32D




Fig. 32E




Fig. 33A




Fig.33B

MAIN

A-V

CPG

SUCTION/FLUIDS

GASES

WAVEFORMS

SETTINGS

PROTOCOL

SENSORS

CPG

TIMERS

PULSE

OTHER

AIR DETECTORS

BUBBLE SENSOR 1

BLENDER/GAS

BLENDER

PRESSURE SENSORS

ARTERIAL LINE

TEMP. SENSORS

VENOUS

LEVEL DETECTORS

CONTINUOUS SENSOR

SAT/HCT

WARNING LIMITS

Fig.33C

AIR DETECTORS

BUBBLE SENSOR 1

BUBBLE SENSOR 1

BUBBLE SENSOR 2

CPG BUBBLE DETECTOR

Fig. 33D




Fig. 33E




Fig. 33F

	PROTOCOL		SENSORS		CPG		TIMERS		PULSE		OTHER			
A-V	BAG LOW WARNING ALARM ON OFF		SELECT & MODIFY BAG PRESETS		BAG 1 PRESET 1		500 ml 100 mEq/L		MODIFY		500 ml 100 mEq/L			
CPG	BAG LOW AUDIO ALARM ON OFF		BAG 2 PRESET 2		500 ml 100 mEq/L		MODIFY		PRESET 3		500 ml 100 mEq/L			
SUCTION/FLUIDS	BAG EMPTY PUMP STOP ON OFF				500 ml 100 mEq/L		MODIFY		PRESET 4		500 ml 100 mEq/L			
GASES	K+ HIGH WARNING ALARM ON OFF													
WAVEFORMS	K+ HIGH AUDIO ALARM ON OFF													
SETTINGS														
CONFIGURE BOLUS			MODE			VOLUME			500			BOLUS COUNT		
												UP DOWN		
SELECT DELIVERY			SELECT BAG FROM WHICH TO DELIVER			1			2			ANTEGRADE		
												RETROGRADE		
												CRYSTALLOID BLOOD/CRYST. ONLY		
												RATIOED		